



Plastic Tray Tags

INTENDED USE:

The Key Surgical® Plastic Tray Tags are intended to be attached to the side of a tray or container for purposes of color coding or affixing a label.

CONTRAINDICATIONS:

There are no known contraindications and/or adverse effects.

INSTRUCTIONS FOR USE:

1. Attach tag to tray by:
 - a. Sliding the tray clip over the edge of the tray, OR:
 - b. Alternatively, some tag designs may require a fastener such as a ring, clip, or tie to connect the tag to the tray.
2. Self-adhesive labels may be affixed to the tag.

POINT OF USE CARE:

Clean plastic tray tag as soon as possible after use. If cleaning must be delayed, immerse in an enzymatic solution or water to prevent drying and encrustation of surgical soil. Remove excessive soil with a disposable wipe.

MANUAL CLEANING:

1. Pre-rinse under cold tap water for one (1) minute to remove gross debris
2. Soak for a minimum of two (2) minutes in a pH neutral detergent, prepared in accordance with the manufacturer's instructions for use.
3. Rinse under cold tap water for one (1) minute.
4. Ultrasonically clean for a minimum of five (5) minutes in a neutral pH detergent, prepared in accordance with the manufacturer's instructions for use.
5. Rinse under cold tap water for one (1) minute.

AUTOMATED CLEANING:

It may be necessary to manually clean prior to automated processing to improve the removal of adherent soil. Follow the previous instructions for manual cleaning.

1. Run the automatic wash cycle – minimum cycle parameters:
 - 1 minute cold pre-rinse
 - 5 minute enzyme wash at 43° C minimum temperature
 - 1 minute cold rinse
 - 7 minute dry at 90° C minimum temperature

CLEANING INSPECTION:

Visually inspect before sterilization or storage to ensure the complete removal of soil from surfaces. If soil is still present, re-clean the plastic tray tag.

STERILIZATION:

Plastic Tray Tags are provided non-sterile and have been validated for functionality in the recommended cycles:

Cycle Type (Steam)	Temperature	Minimum Time – Full Cycle
Gravity	121° C (250° F)	30 minutes
Pre-Vacuum	132° C (270° F)	4 minutes

100% Ethylene Oxide (EtO):

Concentration	735 mg/L
Temperature	55° C
Exposure Time	60 minutes
Humidity	50-70%

STERRAD® System and Cycle within the United States	STERRAD® System and Cycle outside United States
STERRAD® 50	STERRAD® 50
STERRAD® 200	STERRAD® 200 Short cycle
STERRAD® 100S	STERRAD® 100S Short cycle
STERRAD® NX Advanced cycle	STERRAD® NX Advanced cycle
STERRAD® NX Standard cycle	STERRAD® NX Standard cycle

It remains the responsibility of the processor to ensure that the processing, as actually performed using equipment, materials and personnel in the processing facility, achieves the desired result. This requires verification and/or validation and routine monitoring of the process.

DISPOSAL:

Plastic Tray Tags have a life span and will require replacing if there are any signs of deterioration or loss of functionality. Dispose of product following facility policy for disposal.



MANUFACTURER: KEY SURGICAL LLC
8101 WALLACE ROAD / EDEN PRAIRIE, MN 55344 / USA
PH: 800.541.7995 KEYSURGICAL.COM



LOT
NUMBER



CATALOG
NUMBER

BROC-3146, Rev C, 02/2019